10

15

REMARKS

Claims 1-14 are pending in the application. The claims were rejected in the OA as follows.

Claims / Section	35 U.S.C. Sec.	References / Notes
1–3, 5–10, & 12–14	§102(b) Anticipation	 Mangold, et al. (U.S. Patent No. 4,972,487).
4 & 11	§103(a) Obviousness	 Mangold, et al. (U.S. Patent No. 4,972,487); and Levitt, et al. (U.S. Patent No. 4,731,850).

Applicant provides the following discussion distinguishing the present invention from the art cited against it.

35 U.S.C. §102(b), CLAIMS 1-3, 5-10, AND 12-14 ANTICIPATION BY MANGOLD

In the OA, on p. 1, the Examiner rejected claim 1 over Mangold and identified the relevant portions of Mangold asserted as reading on the elements of claim 1. Applicant respectfully traverses this rejection.

Mangold teaches a hearing aid that is capable of changing programs or settings for the signal-processing. This changing can he performed manually or automatically. Each user will require his or her individual number of settings (1:23–34). There are problems if too many different programs are stored in the hearing aid, but only a few of them are used. Therefore, the number of different settings or programs stored In the hearing aid, is limited in Mangold. For such limitation, a dispenser has to know which settings and programs are used by the hearing aid wearer. In order to obtain information about these specific settings and programs, a history of changes in settings is recorded in the hearing aid

10

15

20

(1:40–45). The dispenser then reads out the data stored in the data logging memory, Using this information, a new set of programs can be stored in the memory of the user's device (8:64 – 9:2). This means that Mangold only discloses a recording of setting changes, but does not use this information for automatically controlling the hearing aid. Furthermore, the hearing aid disclosed by Mangold, does not automatically learn the settings in dependence of an actual environment situation.

The Examiner states that Mangold measures at least one sound quantity by "recording/data logging environmental events". However, the section cited by the Examiner describes the recording or logging of a history of changes in setting only. Indeed, the memory permits "recording of environmentally selected events such as selection of settings," etc. This means that, for example, an environmental event initiates an automatic change of a program or setting. However, the actual environment situation is not measured and recorded.

Furthermore, the Examiner states that the step of automatic learning is done by Mangold by "storing events and settings In memory based on recorded events". However, such a step of learning does not take place in the method of Mangold, since in this disclosure, only changes are recorded, but not further processed. The dispenser uses the information "manually" for selecting a new set of programs to be stored in the memory of the hearing aid.

For this reason, Mangold fails to teach or suggest all of the elements of claim 1 and therefore does not anticipate claim 1. A similar argument can be

10

15

20

made for independent claims 7, 13, and 14 as well as dependent claims 2, 3, 5, 6, 8–10, and 12.

35 U.S.C. §103(a), CLAIMS 4 AND 11 OBVIOUSNESS OVER MANGOLD IN VIEW OF LEVITT

In the OA, on pp. 4–5, the Examiner rejected claims 4 and 11 over

Mangold in combination with Levitt, noting that while Mangold does not expressly
disclose the learning ensues via temporal weighting of learning steps, the use of
temporal weighting of input signals is well-known in the art and Levitt teaches
that learning ensues via temporal weight of learning steps (the measured
quantity is continuously monitored and the learning steps are weighted based on
the frequency band and magnitude of the measured quantity, and the level
detector generates a two-bit coefficient of the average signal level to set the
frequency response of the programmable filter in accordance with the changing
environmental situation. The Examiner then noted that it would have been
obvious to one of ordinary skill in the art at the time of the invention to utilize the
temporal weight of Leavitt in the recording of input of Mangold to more accurately
take into account changes in environmental conditions for automatic adjustment
of the hearing aid.

Applicant relies on the discussion above with regard to Mangold and the independent claims of the application, and asserts that the addition of Levitt does not include a teaching for the elements that are not taught by Mangold.

Furthermore, there is no teaching or suggestion for combining the teaching of Levitt with the teaching of Mangold to arrive at the claimed invention. The prior art must suggest the desirability of the claimed invention. (MPEP 2143.01(I). The

temporal learning of claims 4 and 11 must be viewed in the context of the dependent claims from which they depend (claims 1 and 7 respectively), and the contextless addition of the teaching of Levitt is not sufficient to obviate the present reference. The fact that reference can be combined or modified is not sufficient to establish *prima facie* obviousness (MPEP 2143.01(III))

For these reasons, the Applicant asserts that the claim language clearly distinguishes over the cited prior art, and respectfully requests that the Examiner withdraw the §103 rejection from the present application.

CONCLUSION

Inasmuch as each of the objections have been overcome by the amendments, and all of the Examiner's suggestions and requirements have been satisfied, it is respectfully requested that the present application be reconsidered, the rejections be withdrawn and that a timely Notice of Allowance be issued in this case.

15	Respectfully submitted,
	/Mark Bergner/ (Reg. No. 45,877) Mark Bergner SCHIFF HARDIN, LLP
20	PATENT DEPARTMENT
	6600 Sears Tower
	Chicago, Illinois 60606-6473
	(312) 258-5779
	Attorney for Applicants
25	Customer Number 26574